

Environmental Performance using DLP® NIRscan™ Nano

About Test Results

For many applications, performance under extreme or varying environmental conditions must be maintained. Standard requirements for NIR spectrometers include wavelength repeatability to be on the order of 10 percent of the instrument's resolution.¹ Any variability observed could be caused by many parts of the system: variation in the optics, mechanics, or noise differences over temperature. To show stability of the DLP NIRscan Nano evaluation module (EVM), a unit was tested for wavelength variation in a temperature and humidity chamber.

Related Documentation from Texas Instruments

- DLPC150 Data Sheet: *DLPC150 DLP® Digital Controller for Advanced Light Control*, TI literature number [DLPS048](#)
- DLP2010NIR Data Sheet: *DLP2010NIR (0.2 WVGA Near-Infrared DMD)*, TI literature number [DLPS059](#)
- Programmer's Guide: *DLPC150 Programmer's Guide*, TI literature number [DLPU031](#)
- Application Note: *Texas Instruments DLP® Spectrometer Design Considerations*, TI literature number [DLPA049](#)
- Application Note: *Flexible Trade-offs in Maximizing SNR and Resolution in TI DLP® Technology-Based Spectrometer Systems*, TI literature number [DLPA066](#)
- Selection Guide: *TI DLP® Technology for Spectroscopy*, TI literature number [DLPT020](#)

If You Need Assistance

Refer to the [DLP Products and MEMS TI E2E Community support forums](#)

Experimental Setup

An argon low-pressure discharge wavelength calibration lamp whose spectrum is shown in Figure 1 was placed outside the temperature and humidity chamber, and fiber coupled into the spectrometer which was placed inside the controlled chamber. Three separate experiments were run. In the first experiment, the temperature was varied while the dew point was held constant at 9.5°C. Similarly, the second experiment varied the temperature while holding the dew point constant at 18°C. In the third experiment, the relative humidity was held constant while varying the temperature, such that the absolute humidity was directly proportional to temperature. A diagram of the experimental setup can be seen in Figure 2 below.




Figure 1: The spectrum of the argon test source




Figure 2: Experimental setup diagram

For each of the three experiments, the argon spectrum was scanned multiple times. The first scan was used as the reference values for the six peak wavelengths. The subsequent scans were measured at the different environmental conditions and used to calculate the variations from the reference values. To abide by the 10 percent standard presented earlier, the variation needs to be on the order of $\pm 0.8\text{nm}$ or less.

		Reference			Test			Variation		
		T_1			T_2	T_3	T_4	V_1	V_2	V_3
Wavelength Peaks	λ_A	M_{A1}	M_{A2}	M_{A3}	M_{A4}	$M_{A2}-M_{A1}$	$M_{A3}-M_{A1}$	$M_{A4}-M_{A1}$		
	λ_B	M_{B1}	M_{B2}	M_{B3}	M_{B4}	$M_{B2}-M_{B1}$	$M_{B3}-M_{B1}$	$M_{B4}-M_{B1}$		
	λ_C	M_{C1}	M_{C2}	M_{C3}	M_{C4}	$M_{C2}-M_{C1}$	$M_{C3}-M_{C1}$	$M_{C4}-M_{C1}$		
	λ_D	M_{D1}	M_{D2}	M_{D3}	M_{D4}	$M_{D2}-M_{D1}$	$M_{D3}-M_{D1}$	$M_{D4}-M_{D1}$		
	λ_E	M_{E1}	M_{E2}	M_{E3}	M_{E4}	$M_{E2}-M_{E1}$	$M_{E3}-M_{E1}$	$M_{E4}-M_{E1}$		
	λ_F	M_{F1}	M_{F2}	M_{F3}	M_{F4}	$M_{F2}-M_{F1}$	$M_{F3}-M_{F1}$	$M_{F4}-M_{F1}$		

Figure 3: Example measurements for each experiment

Performance Results

Two experiments were conducted at constant dew points of 9.5°C and 18°C. Because the dew points were held constant, this tested the wavelength stability over temperature independent of water vapor content of the air. The distribution of the peak locations are shown in Figures 4 and 5. The variations were observed to be much better than the $\leq \pm 0.8\text{nm}$ target.

The third experiment was conducted where relative humidity was held constant and the temperature was varied over a wider range. In this series, the absolute humidity was increasing along with the temperature at each data point. The results are shown in Figure 6, with the variations again well within the $\leq \pm 0.8\text{nm}$ goal.




Figure 2: Wavelength location variation vs temperature at 9.5°C dew point




Figure 3: Wavelength location variation vs temperature at 18°C dew point




Figure 4: Wavelength location variation vs temperature with 60% relative humidity

Conclusion

The CMOS based, landed MEMS digital micromirror device (DMD) architecture in the DLP NIRscan Nano EVM has been shown to provide stable wavelength accuracy in this compact system over the wide range of temperatures and humidities tested here. NIR spectrometer applications which have previously been hampered from the market by high cost or questionable field performance now have a platform available that combines the best in handheld, portable performance.

REFERENCES

- [1] Haibach, F. G., “Getting the best results from nir measurements,” tech. rep., BaySpec Incorporated (2014).

IMPORTANT NOTICE AND DISCLAIMER

TI PROVIDES TECHNICAL AND RELIABILITY DATA (INCLUDING DATA SHEETS), DESIGN RESOURCES (INCLUDING REFERENCE DESIGNS), APPLICATION OR OTHER DESIGN ADVICE, WEB TOOLS, SAFETY INFORMATION, AND OTHER RESOURCES "AS IS" AND WITH ALL FAULTS, AND DISCLAIMS ALL WARRANTIES, EXPRESS AND IMPLIED, INCLUDING WITHOUT LIMITATION ANY IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR A PARTICULAR PURPOSE OR NON-INFRINGEMENT OF THIRD PARTY INTELLECTUAL PROPERTY RIGHTS.

These resources are intended for skilled developers designing with TI products. You are solely responsible for (1) selecting the appropriate TI products for your application, (2) designing, validating and testing your application, and (3) ensuring your application meets applicable standards, and any other safety, security, regulatory or other requirements.

These resources are subject to change without notice. TI grants you permission to use these resources only for development of an application that uses the TI products described in the resource. Other reproduction and display of these resources is prohibited. No license is granted to any other TI intellectual property right or to any third party intellectual property right. TI disclaims responsibility for, and you will fully indemnify TI and its representatives against, any claims, damages, costs, losses, and liabilities arising out of your use of these resources.

TI's products are provided subject to [TI's Terms of Sale](#) or other applicable terms available either on ti.com or provided in conjunction with such TI products. TI's provision of these resources does not expand or otherwise alter TI's applicable warranties or warranty disclaimers for TI products.

TI objects to and rejects any additional or different terms you may have proposed.

Mailing Address: Texas Instruments, Post Office Box 655303, Dallas, Texas 75265
Copyright © 2021, Texas Instruments Incorporated